



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/669,887	09/24/2003	Xuefeng Cheng	PHOTP003	2242

26541 7590 07/25/2005

Cindy S. Kaplan  
P.O. BOX 2448  
SARATOGA, CA 95070

EXAMINER

WINAKUR, ERIC FRANK

ART UNIT PAPER NUMBER

3736

DATE MAILED: 07/25/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

<b>Office Action Summary</b>	Application No. 10/669,887	Applicant(s) CHENG ET AL	
	Examiner Eric F. Winakur	Art Unit 3736	

**-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --**  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☐ Responsive to communication(s) filed on \_\_\_\_.
- 2a) ☐ This action is **FINAL**.                      2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-17 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-17 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All    b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_.
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- |   |  |
|---|--|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892)   | 4) <input type="checkbox"/> Interview Summary (PTO-413)<br>Paper No(s)/Mail Date. ____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948)                                  | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152)            |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)<br>Paper No(s)/Mail Date ____ | 6) <input type="checkbox"/> Other: ____  |

## **DETAILED ACTION**

### ***Claim Objections***

1. Claim 16 is objected to because of the following informalities: Claim 16 improperly depends from itself. It appears that it should depend from claim 15 and for the purpose of examination, will be considered to depend therefrom. Appropriate correction is required.

### ***Claim Rejections - 35 USC § 102***

2. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

3. Claims 1 - 5, 8, 9, and 11 - 14 are rejected under 35 U.S.C. 102(e) as being anticipated by Hueber. Hueber teaches an optical measurement system that includes at least two emitters and at least two detectors arranged to perform measurements at multiple wavelengths for determining absolute concentration of hemoglobin molecules or other analytes (column 3, lines 17 - 33). Calculations are based on wave equations, as discussed in column 5, line 61 - column 8, line 9. Further details are provided in the section on the Self-Referencing Absolute Probe that begins in column 10.

### ***Claim Rejections - 35 USC § 103***

Art Unit: 3736

4. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

5. Claims 6, 7, 10, and 15 - 17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Hueber, as applied to claims 1, 8, and 12 above. Hueber teaches an apparatus and method for measuring absolute concentrations of analytes, but does not particularly teach that the device is used in a surgical setting. Further, Hueber teach the mathematical basis of their measurement, but do not teach the particular form of the equation set forth in claims 15 - 17. With regard to use of the device of Hueber in a surgical setting, it is well known in the art that physiological monitoring of surgical patients and surgical tissue is performed to monitor patient/tissue status during such a procedure. Thus, it would have been obvious to one of ordinary skill in the art at the time of the invention to implement Hueber in a surgical setting, since this is a well-known setting to use Hueber, and it is generally held to be within the skill level of the art to use a device for a known purpose. With regard to different forms of the underlying equations used by both Applicant and Hueber, it would have been within the skill level of the art at the time of the invention to arrange and optimize the mathematical equations of Hueber to account for particular sensor/detector/processor arrangements, including those used and set forth in Applicant's claims.

***Double Patenting***

Art Unit: 3736

6. The nonstatutory double patenting rejection is based on a judicially created doctrine grounded in public policy (a policy reflected in the statute) so as to prevent the unjustified or improper timewise extension of the "right to exclude" granted by a patent and to prevent possible harassment by multiple assignees. See *In re Goodman*, 11 F.3d 1046, 29 USPQ2d 2010 (Fed. Cir. 1993); *In re Longi*, 759 F.2d 887, 225 USPQ 645 (Fed. Cir. 1985); *In re Van Ornum*, 686 F.2d 937, 214 USPQ 761 (CCPA 1982); *In re Vogel*, 422 F.2d 438, 164 USPQ 619 (CCPA 1970); and, *In re Thorington*, 418 F.2d 528, 163 USPQ 644 (CCPA 1969).

A timely filed terminal disclaimer in compliance with 37 CFR 1.321(c) may be used to overcome an actual or provisional rejection based on a nonstatutory double patenting ground provided the conflicting application or patent is shown to be commonly owned with this application. See 37 CFR 1.130(b).

Effective January 1, 1994, a registered attorney or agent of record may sign a terminal disclaimer. A terminal disclaimer signed by the assignee must fully comply with 37 CFR 3.73(b).

7. Claim 1 is rejected under the judicially created doctrine of obviousness-type double patenting as being unpatentable over claim 26 of U.S. Patent No. 6,587,703. Although the conflicting claims are not identical, they are not patentably distinct from each other because the claims of the instant application are broader than those of the patent. Thus, any method meeting the steps of the patent would necessarily meet those of the instant application.

### **Conclusion**


8. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Hayashi (Figures 4, 8, 10; specification column 4 - 5) teaches an instrument for measuring concentrations of constituent parts of blood that includes a probe emitting first and second wavelengths, a detector for measuring light that has interacted with a subject's tissue, and a calculator for determining the concentrations.

Art Unit: 3736

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Eric F. Winakur whose telephone number is 571/272-4736. The examiner can normally be reached on M-Th, 7:30-5; alternate Fri.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Max Hindenburg can be reached on 571/272-4726. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Eric F Winakur  
Primary Examiner  
Art Unit 3736

22 July 2005